



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

GENERAL FASTENERS COMPANY
1524 East 14 Mile Road
Madison Heights, MI 48071
Greg Kaske Phone: 734 452 2400

MECHANICAL

Valid to: July 31, 2026

Certificate Number: 0170.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following types of metals and metal products tests (including fasteners and automotive components):

<u>Test:</u>	<u>Test Method(s) ¹:</u>
Hardness (Rockwell A, B, C, 15N, 30N)	ASTM E18, F606/F606M; ASME B18.13; SAE J81, J417, J933
Microhardness (Knoop, Vickers: 300g, 500g)	ASTM E384; SAE J81, J933
Tension (Axial, Wedge: 120,000 lb. Capacity)	ASTM F606/F606M
Proof Load (Internally & Externally Threaded)	ASTM F606/F606M; IFI 100/107 (2007); ASME B18.16.6
Stress Durability (Hydrogen Embrittlement)	ASME B18.6.3, B18.13, B18.21.1; SAE J81, J773, J1237; SAE/USCAR-7
Torsion	ASTM F606/F606M; ASME B18.6.3; SAE J81, J933, J1237
Ductility	ASME B18.6.3; SAE J81, J1237
Salt Spray	ASTM B117
Drive Test	ASME B18.6.3; SAE J81, J933, J1237
Twist Test	ASME B18.13; ASME B18.21.1
X-ray Fluorescence of Plating Thickness	ASTM B568
<u>Metallographic Evaluation</u>	
Case Depth	SAE J423
Depth of Decarburization	ASTM F2328, F2328M; GM6104M (2011); SAE J419; ISO 898-1, 898-5 (visual only)
Surface Discontinuities	ASTM A574, A574M, F788, F812, F835, F835M, F912, F912M; SAE J123; ISO 6157-1, 6157-2, 6157-3; GM6102M (2011)

(A2LA Cert. No. 0170.01) Revised 07/22/2024

Page 1 of 3

Test:**Test Method(s) ¹:****Metallographic Evaluation (continued)**

Microstructure

ASM Handbook, Volume 9

SEM/EDS

ASTM E1508

Fracture Analysis

ASM Handbook, Volume 11 and the methods listed above

Chemical

Optical Emission of Carbon, Alloy & Stainless Steel

ASTM E415, E1086

(Al, B, C, Cr, Cu, Fe, Mn, Mo, Ni, P, S, Si, V)

II. Dimensional Testing ²

Parameter/Equipment	Range	CMC ³ (±)	Comments
Linear ⁴ – 1D	Up to 0.070 in (50X) Up to 0.070 in (100X) Up to 12 in Up to 2 in Up to 10 in	0.0009 in 0.0009 in 0.0005 in 0.0003 in 0.0005 in	Metallograph / Image analysis system / MIL-STD-120 (1996) Calipers / MIL-STD-120 (1996) Micrometer / MIL-STD-120 (1996) Height gage / MIL-STD-120 (1996)
2D	X Axis: 10 in Y Axis: 6 in	0.0002 in 0.0002 in	Optical comparator / MIL-STD-120 (1996)
Angle ⁴	0° to 360°	4'	Optical comparator / MIL-STD-120 (1996)
Radius ⁴	(0.005 to 5) in	0.0002 in	Optical comparator / MIL-STD-120 (1996)
Threads ⁴ –			
External	(0-80 to 2.875-12) in M3x0.5 to M30x1.5	N/A	Ring gage / ASME B1.3; MIL-STD-120 (1996) system 21
	(6-32 to 1-14) in M8x1.25 to M16x2.0	0.0002 in	Tri-Roll Pitch micrometer / ASME B1.3; MIL-STD-120 (1996) system 22
Internal	(2-56 to 1½ -12) in M3x0.5 to M48x2.0	N/A	Plug gages / ASME B1.3; MIL-STD-120 (1996) system 21

¹ When the date, edition, version, etc. is not identified in the scope of accreditation, laboratories may use the version that immediately precedes the current version for a period of one year from the date of publication of the standard measurement method, per part C., Section 1 of A2LA *R101 - General Requirements- Accreditation of ISO-IEC 17025 Laboratories*.

² Commercial dimensional testing service is sometimes available for this laboratory.

³ Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine measurements of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific measurement performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific measurement.

⁴ This test is not equivalent to that of a calibration.



Accredited Laboratory

A2LA has accredited

GENERAL FASTENERS COMPANY

Madison Heights, MI

for technical competence in the field of

Mechanical Testing

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 19th day of April 2024.

A blue ink signature of Mr. Trace McInturff, written over a horizontal line.

Mr. Trace McInturff, Vice President, Accreditation Services
For the Accreditation Council
Certificate Number 0170.01
Valid to July 31, 2026
Revised July 22, 2024

For the types of tests to which this accreditation applies, please refer to the laboratory's Mechanical Scope of Accreditation.